

**RUO ELISA reaģentu saraksts 2016 (tikai Latvijas teritorijā)**

Kat.Nr.	Produkts	Testu skaits
<b>HIV</b>		
D-0134	<b>HIV-1 p24 – antigen – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection and confirmation of HIV-1 p24 antigen presence.	
D-0151	<b>CombiBest HIV - 1,2 AG/AB (set 1)</b>	24x8
D-0152	<b>CombiBest HIV - 1,2 AG/AB (set 2)</b>	12x8
	Enzyme immunoassay kit for the detection of anti-HIV - 1,2 antibodies and HIV-1 p24 antigen.	
D-0193	<b>UniBest HIV – 1,2 Ab (set 1)</b>	24x8
D-0192	<b>UniBest HIV - 1,2 Ab (set 2)</b>	12x8
	Enzyme immunoassay kit for the detection of total antibodies against HIV-1,2.	
D-0160	<b>LIA – blot - HIV - 1,2 BEST</b>	24
	Reagent kit for the detection of antibodies against HIV- 1,2 antigens by immunoblot.	
<b>HEPATITIS A</b>		
D-0352	<b>Vectohep A - IgM</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against hepatitis A virus.	
D-0356	<b>HAV – antigen – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of hepatitis A virus antigen.	
D-0362	<b>Vectohep A - IgG</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against hepatitis A virus.	
<b>HEPATITIS B</b>		
D-0543	<b>HBsAg – EIA – BEST (set 2)</b>	192
D-0544	<b>HBsAg – EIA – BEST (set 3)</b>	12x8
	Enzyme immunoassay kit for the detection of HBs-antigen. Sensitivity – 0.01 ng/ml.	
D-0545	<b>HBsAg – EIA – BEST (quantitative analysis)</b>	12 x 8
	Enzyme immunoassay kit for the quantitative analysis of HBs-antigen. Sensitivity – 0.05 ng/ml. Measurement range: 0-10ng/ml.	

	<b>HBsAg – confirmatory – EIA - BEST</b>	
D-0546	Enzyme immunoassay kit for the confirmation of HBs-antigen presence. Sensitivity – 0.01 ng/ml.	6x8
D-0557	<b>Vectohep B – HBs – antigen (set 1)</b>	24x8
D-0556	<b>Vectohep B – HBs – antigen (set 3)</b>	12x8
	Enzyme immunoassay kit for the detection of HBs-antigen. Sensitivity – 0.05 ng/ml.	
D-0558	<b>Vectohep B-HBs – antigen (set 5)</b> Enzyme immunoassay kit for the detection and confirmation of HBs-antigen presence. Sensitivity – 0.05 mg/ml.	6x8
D-0562	<b>VectoHBsAg - antibodies</b> Enzyme immunoassay kit for the qualitative and quantitative determination of antibodies against HBs-antigen	12x8
D-0566	<b>VectoHBcAg - antibodies</b> Enzyme immunoassay kit for the detection of total antibodies against hepatitis B core- antigen.	12x8
D-0574	<b>HepaBest anti – HBc - IgG</b> Enzyme immunoassay kit for the detection of IgG against hepatitis B core- antigen.	12x8
D-0564	<b>VectoHBcAg - IgM</b> Enzyme immunoassay kit for the detection of IgM against hepatitis B core-antigen.	12x8
D-0576	<b>VectoHBe - antigen</b> Enzyme immunoassay kit for the detection of E-antigen of the hepatitis B virus.	12x8
D-0578	<b>VectoHBe - IgG</b> Enzyme immunoassay kit for the detection of IgG against HBe-antigen of the hepatitis B virus.	12x8
	<b>HEPATITIS C</b>	
D-0760	<b>RecombiBest anti-HCV-IgM</b> Enzyme immunoassay kit for the detection of IgM against hepatitis C virus.	12x8
D-0773	<b>Best anti-HCV (set 1/strip)</b>	24x8
D-0772	<b>Best anti-HCV (set 2)</b>	12x8
	Enzyme immunoassay kit for the detection of IgG and IgM against hepatitis C virus.	
D-0774	<b>Best anti-HCV – SPECTR</b> Enzyme immunoassay kit for the detection of antibodies to individual proteins of hepatitis C virus (core, NS3, NS4 NS5)	6x4
	<b>Best anti-HCV (set 4)</b>	

D-0776	Enzyme immunoassay kit for the detection and confirmation of IgG and IgM against hepatitis C virus.	12x8
D-0777	<b>HCV AG/AB – IEA – BEST (set 1)</b>	12x8
D-0778	<b>HCV AG/AB – IEA – BEST (set 2)</b>	24x8
	Enzyme immunoassay kit for the detection of core-antigen against HCV and antibodies to hepatitis C virus.	
	<b>HCV core-antigen-IEA – BEST</b>	
D-0779	Enzyme immunoassay kit for the detection of core-antigen against hepatitis C virus.	12x8
<b>HEPATITIS D</b>		
	<b>Vectohep D - IgM</b>	
D-0952	Enzyme immunoassay kit for the detection of IgM against hepatitis D virus.	12x8
	<b>Vectohep D - antibodies</b>	
D-0954	Enzyme immunoassay kit for the detection of total antibodies against hepatitis D virus.	12x8
<b>HEPATITIS E</b>		
	<b>Vectohep E - IgG</b>	
D-1056	Enzyme immunoassay kit for the detection of IgG against hepatitis E virus.	12x8
	<b>Vectohep E - IgM</b>	
D-1058	Enzyme immunoassay kit for the detection of IgM against hepatitis E virus.	12x8
<b>VIRAL INTESTINAL INFECTIONS</b>		
	<b>Rotavirus – antigen – EIA - BEST</b>	
D-1652	Enzyme immunoassay kit for the detection of human Rotavirus antigen.	12x8
	<b>Adenovirus – antigen – EIA - BEST</b>	
D-1654	Enzyme immunoassay kit for the detection of human Adenovirus antigen	12x8
<b>NOROVIRUS</b>		
	<b>Norovirus – antigen – EIA - BEST</b>	
D-1656	Enzyme immunoassay kit for the detection of Norovirus antigen genogroup I and II in feces.	12x8
<b>TICK-BORNE ENCEPHALITIS</b>		
	<b>VectoTBE - IgM</b>	
D-1152	Enzyme immunoassay kit for the detection of IgM against tick-borne encephalitis virus.	12x8
	<b>VectoTBE - antigen</b>	
D-1154	Enzyme immunoassay kit for the detection of tick-borne encephalitis virus antigen.	12x8

D-1156	<b>VectoTBE - IgG</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against tick-borne encephalitis virus.	
<b>BORRELIOSIS</b>		
D-1452	<b>LymeBest – IgG</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against causative agent of tick-borne borreliosis ( <i>B. afzelii</i> , <i>B. garinii</i> , <i>B. burgdorferi s.s.</i> ).	
D-1454	<b>LymeBest – IgM</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against causative agent of tick-borne borreliosis ( <i>B. afzelii</i> , <i>B. garinii</i> , <i>B. burgdorferi s.s.</i> ).	
<b>HEMORRHAGIC FEVER WITH RENAL SYNDROME (HFRS)</b>		
D-4902	<b>VectoHanta – IgG</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against hantavirus.	
D- 4904	<b>VectoHanta – IgM</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against hantavirus.	
<b>CRIMEAN-CONGO HAEMORRHAGIC FEVER</b>		
D-5052	<b>VectoCrimea – CHF - IgG</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against Crimean-Congo hemorrhagic fever virus.	
D-5054	<b>VectoCrimea – CHF - IgM</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against Crimean-Congo hemorrhagic fever virus.	
D-5056	<b>VectoCrimea – CHF - antigen</b>	12x8
	Enzyme immunoassay kit for the detection of Crimean-Congo hemorrhagic fever virus antigen.	
<b>WEST NILE VIRUS (WNV)</b>		
D-5150	<b>VectoNile - IgM</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against West Nile virus.	
D-5152	<b>VectoNile - IgG</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against West Nile virus.	
D-5154	<b>VectoNile – IgG - avidity</b>	6x8
	Enzyme immunoassay kit for the determination of avidity index of IgG against West Nile virus.	
<b>MEASLES</b>		
D-1356	<b>VectoMeasles - IgG</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against Measles virus.	
	<b>VectoMeasles - IgM</b>	

D-1358	Enzyme immunoassay kit for the detection of IgM against Measles virus.	12x8
<b>MUMPS</b>		
D-2602	<b>VectoMumps - IgG</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against Mumps virus.	
D-2604	<b>VectoMumps - IgM</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against Mumps virus.	
<b>CYTOMEGALOVIRUS</b>		
D-1552	<b>VectoCMV - IgM</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against Cytomegalovirus.	
D-1556	<b>CMV – IgG – EIA - BEST</b> Enzyme immunoassay kit for the quantitative and quantitative determination of IgG against Cytomegalovirus.	12x8
D-1560	<b>CMV – IgG – blot - BEST</b>	20
	Reagent kit for the detection of IgG against Cytomegalovirus by immunoblot.	
D-1558	<b>VectoCMV – IgG - avidity</b>	6x8
	Enzyme immunoassay kit for the determination of avidity index of IgG against Cytomegalovirus.	
D-1566	<b>VectoCMV – IEA - antibodies</b>	12x8
	Enzyme immunoassay kit for the detection of IgG and IgM against Cytomegalovirus immediate early antigen.	
<b>TOXOPLASMA</b>		
D-1752	<b>VectoToxo - IgG</b>	12x8
	Enzyme immunoassay kit for the quantitative and quantitative determination of IgG against <i>Toxoplasma gondii</i> .	
D-1754	<b>VectoToxo - IgA</b>	12x8
	Enzyme immunoassay kit for the detection of IgA against <i>Toxoplasma gondii</i> .	
D-1756	<b>VectoToxo - IgM</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against <i>Toxoplasma gondii</i> .	
D-1760	<b>Toxo – IgM- IEA-BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against <i>Toxoplasma gondii</i> (Antibody capture method)	
D-1762	<b>VectoToxo – IgG - avidity</b>	6x8
	Enzyme immunoassay kit for the determination of avidity index of IgG against <i>Toxoplasma gondii</i> .	
	<b>VectoToxo - antibodies</b>	

D-1764	Enzyme immunoassay kit for the detection of total antibodies against <i>Toxoplasma gondii</i> in human or animal serum or plasma	12x8
<b>SYPHILIS</b>		
D-1851	<b>RecombiBest antipallidum – IgG (set 1)</b>	96x2
D-1852	<b>RecombiBest antipallidum – IgG (set 2)</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Treponema pallidum</i> .	
D-1857	<b>RecombiBest antipallidum - total antibodies (set 1)</b>	24x8
D-1856	<b>RecombiBest antipallidum - total antibodies (set 2)</b>	12x8
	Enzyme immunoassay kit for the detection of total antibodies against <i>Treponema pallidum</i> .	
D-1858	<b>RecombiBest antipallidum - IgM</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against <i>Treponema pallidum</i> .	
D-1860	<b>Treponema pallidum - blot - BEST</b>	24
	Reagent kit for the detection of antibodies against <i>Treponema pallidum</i> by immunoblot.	
D-1820	<b>RPHA - BEST - antipallidum</b>	100
	Reagent kit for the detection of antibodies against <i>Treponema pallidum</i> by passive hemagglutination test.	
<b>CHLAMYDIOSIS</b>		
D-1944	<b>Chlamydophila pneumoniae – IgG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Chlamydophila pneumoniae</i> .	
D-1946	<b>Chlamydophila pneumoniae – IgM – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against <i>Chlamydophila pneumoniae</i> .	
D-1948	<b>Chlamydophila pneumoniae – IgA – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgA against <i>Chlamydophila pneumoniae</i> .	
D-1964	<b>ChlamyBest C.trachomatis - IgG</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Chlamydia trachomatis</i> .	
D-1966	<b>ChlamyBest C.trachomatis - IgM</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against <i>Chlamydia trachomatis</i> .	
D-1968	<b>ChlamyBest C.trachomatis - IgA</b>	12x8
	Enzyme immunoassay kit for the detection of IgA against <i>Chlamydia trachomatis</i> .	
D-1972	<b>ChlamyBest cHSP60 - IgG</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Chlamydia trachomatis</i> heat shock protein (cHSP60).	
D-1982	<b>ChlamyBest MOMP + pgp3 - IgG</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Chlamydia trachomatis</i> major outer membrane protein	

## TRICHOMONIASIS

D-2052	<b>Trichomonas vaginalis - IgG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Trichomonas vaginalis</i> .	
D-2058	<b>Trichomonas vaginalis – IgA – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgA against <i>Trichomonas vaginalis</i> .	

## HERPES VIRUSES

D-2152	<b>VectoHSV - IgG</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against herpes simplex virus.	
D-2156	<b>VectoHSV - 1,2 – IgG - avidity</b>	6x8
	Enzyme immunoassay kit for the determination of avidity index of IgG against HSV-1, 2.	
D-2154	<b>VectoHSV - IgM</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against HSV-1, 2.	
D-2180	<b>VectoHSV–2 - IgG</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against HSV-2.	
D-2182	<b>VectoHSV-2 – IgG - avidity</b>	6x8
	Enzyme immunoassay kit for the determination of avidity index of IgG against HSV-2.	
D-2181	<b>VectoHSV–2 - IgM</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against HSV-2.	
D-2192	<b>VectoVZV - IgG</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against Varicella-Zoster virus.	
D-2188	<b>VectoVZV - IgM</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against Varicella-Zoster virus.	
D-2186	<b>VectoVZV – gE - IgG</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against glycoprotein E of Varicella-Zoster virus.	
D-2184	<b>VectoEBV – VCA - IgG</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against viral capsid antigen of Epstein-Barr virus.	
D-2183	<b>VectoEBV–VCA–IgG –avidity</b>	6x8
	Enzyme immunoassay kit for the determination of IgG avidity index against VCA of EBV.	
D-2176	<b>VectoEBV – VCA - IgM</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against viral capsid antigen of Epstein-Barr virus.	

D-2170	<b>VectoEBV – NA - IgG</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against nuclear antigen (NA) of Epstein-Barr virus.	
D-2172	<b>VectoEBV – EA - IgG</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against early antigen (EA) of Epstein-Barr virus.	
D-2166	<b>VectoHHV-6 - IgG</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against HHV-6.	
D-2160	<b>VectoHHV-8 – IgG</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against HHV-8.	
<b>UREAPLASMOSIS</b>		
D-2254	<b><i>Ureaplasma urealyticum</i> – IgG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Ureaplasma urealyticum</i> antigens.	
D-2258	<b><i>Ureaplasma urealyticum</i> – IgA – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgA against <i>Ureaplasma urealyticum</i> antigens.	
<b>MYCOPLASMOSIS</b>		
D-4352	<b><i>Mycoplasma hominis</i> – IgG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Mycoplasma hominis</i> .	
D-4358	<b><i>Mycoplasma hominis</i> – IgA – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgA against <i>Mycoplasma hominis</i> .	
D-4362	<b><i>Mycoplasma pneumonia</i> – IgG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Mycoplasma pneumoniae</i> .	
D-4366	<b><i>Mycoplasma pneumonia</i> – IgM – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against <i>Mycoplasma pneumoniae</i> .	
D-4368	<b><i>Mycoplasma pneumonia</i> – IgA – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgA against <i>Mycoplasma pneumoniae</i> .	
<b>TUBERCULOSIS</b>		
D-2352	<b>AB – Tub - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of total antibodies against <i>Mycobacterium tuberculosis</i> .	
<b>BRUCELLOSIS</b>		
D-3652	<b>Brucellosis – IgG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Brucellosis</i> .	
D-3654	<b>Brucellosis – IgA – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgA against <i>Brucellosis</i> .	



D-3656	<b>Brucellosis – IgM – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against <i>Brucellosis</i> .	
D-3658	<b>Brucellosis – antibodies – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of total antibodies against <i>Brucellosis</i> .	
<b>YERSINIOSIS</b>		
D-3202	<b>Yersinia – IgG – EIA –BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Yersinia</i> .	
D-3204	<b>Yersinia – IgA – EIA –BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgA against <i>Yersinia</i> .	
D-3206	<b>Yersinia – IgM – EIA –BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Yersinia</i> .	
<b>RUBELLA</b>		
D-2552	<b>VectoRubella - IgG</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against Rubella virus.	
D-2554	<b>VectoRubella - IgM</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against Rubella virus.	
D-2556	<b>VectoRubella – IgG - avidity</b>	6x8
	Enzyme immunoassay kit for the determination of avidity index of IgG against Rubella virus.	
D-2560	<b>Rubella – IgM-EIA-BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgM to Rubella virus by «capture» method.	
<b>PARASITIC DISEASES</b>		
D-2752	<b>Toxocara – IgG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Toxocara canis</i> antigens.	
D-2952	<b>Opisthorchis – IgG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Opisthorchis felineus</i> antigens.	
D-2954	<b>Opisthorchis – IgM – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against <i>Opisthorchis felineus</i> antigens.	
D-2956	<b>Opisthorchis – CIC – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of specific circulatory immune complexes (CIC), containing	
D-2958	<b>Clonorchis – IgG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Clonorchis sinensis</i> antigens.	
D-3152	<b>Trichinella – IgG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Trichinella spiralis</i> antigens.	

D-3154	<b>Trichinella – IgM – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against <i>Trichinella spiralis</i> antigens.	
D-3354	<b>Helminths – IgG – EIA – BEST</b>	3x64
	Enzyme immunoassay kit for the detection of IgG against <i>Opisthorchis felineus</i> , <i>Trichinella</i> , <i>Toxocara</i> and	
D-3356	<b>Echinococcus – IgG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Echinococcus granulosus</i> antigens.	
D-3452	<b>Ascaris – IgG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Ascaris lumbricoides</i> antigens.	
D-3454	<b>Anisakis-IgG-EIA-BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Anisakis nematodes</i> antigens.	
D-3456	<b>Cysticercus-IgG-EIA-BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Taenia solium</i> antigens.	
D-3552	<b>Lamblia – antibodies – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgA, IgM and IgG against <i>Giardia lamblia</i> antigens.	
D-3554	<b>Lamblia – IgM – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against <i>Giardia lamblia</i> antigens.	
D-3556	<b>Lamblia – antigen – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of <i>Giardia lamblia</i> antigens in faeces.	
<b>GASTROINTESTINAL DISEASES</b>		
D-3752	<b>HelicoBest - antibodies</b>	12x8
	Enzyme immunoassay kit for the detection of total antibodies against CagA antigen of <i>Helicobacter pylori</i> .	
D-3754	<b>IgG – Gliadin – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of IgG against gliadin. Measurement range: 0-100 IU/ml. Sensitivity: 2.5 IU/ml.	
D-3756	<b>IgA – Gliadin – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of IgA against gliadin. Measurement range: 0-100 IU/ml. Sensitivity: 2.5 IU/ml.	
D-3758	<b>IgA – transglutaminase – EIA - BEST</b>	6x8
	Enzyme immunoassay kit for the quantitative determination of IgA auto-antibodies against tissue	
	Measurement range: 2.5-100 IU/ml. Sensitivity: 2.5 IU/ml.	
D-3760	<b>IgG – transglutaminase – EIA - BEST</b>	6x8
	Enzyme immunoassay kit for the quantitative determination of IgG auto-antibodies against tissue	

D-3762	<b>Pepsinogen I – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of Pepsinogen I.	
	Measurement range: 0-200 mkg/l.	
	Sensitivity: 1 mkg/l.	
D-3764	<b>Pepsinogen II – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of Pepsinogen II.	
	Measurement range: 0-50 mkg/l.	
	Sensitivity: 1 mkg/l.	
<b>HORMONES</b>		
X-3952	<b>TSH – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of thyroid stimulation hormone (TSH).	
	Measurement range: 0-16 mIU/l.	
	Sensitivity: 0.1 mIU/l.	
X-3954	<b>T<sub>3</sub> total – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of total triiodothyronine.	
	Measurement range: 0-9 nmol/l.	
	Sensitivity: 0.2 nmol/l.	
X-3970	<b>T<sub>3</sub> free – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of free triiodothyronine.	
	Measurement range: 0-20 pmol/l.	
	Sensitivity: 0.5 pmol/l.	
X-3956	<b>T<sub>4</sub> total – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of total thyroxin.	
	Measurement range: 0-400 nmol/l.	
	Sensitivity: 10 nmol/l.	
X-3962	<b>T<sub>4</sub> free – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of free thyroxin.	
	Measurement range: 0-80 pmol/l.	
	Sensitivity: 1 pmol/l.	
X-3966	<b>TG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of thyroglobulin.	
X-3958	<b>Anti-TG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of total antibodies against thyroglobulin.	
	Measurement range: 0-2000 IU/ml.	
	Sensitivity: 10 IU/ml.	
X-3968	<b>Anti – TPO – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of total antibodies against thyroid	
	Measurement range: 0-1000 IU/ml.	
	Sensitivity: 5 IU/ml.	

X-3960	<b>Prolactin – EIA – BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of prolactin.	
	Measurement range: 0-3000 mIU/l.	
	Sensitivity: 15 mIU/l.	
X-3964	<b>Cortisol – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of cortisol.	
	Measurement range: 0-1200 nmol/l	
	Sensitivity: 5 nmol/l.	
X-3972	<b>Testosterone – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of testosterone.	
	Measurement range: 0-60 nmol/l.	
	Sensitivity: 0.2 nmol/l.	
X-3974	<b>FSH – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of follicle-stimulating hormone.	
	Measurement range: 0-100 mIU/ml.	
	Sensitivity: 1 mIU/ml.	
X-3976	<b>LH – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of luteinizing hormone.	
	Measurement range: 0-100 mIU/ml.	
	Sensitivity: 0.5 mIU/ml.	
X-3978	<b>Progesterone – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of progesterone.	
	Measurement range: 0-100 nmol/l.	
	Sensitivity: 0.4 nmol/l.	
<b>PREGNANCY</b>		
D-4154	<b>HCG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of human chorionic gonadotropin.	
	Measurement range: 0-500 MIU/l.	
	Sensitivity: 2 MIU/l.	
D-4158	<b>Free <math>\beta</math>-HCG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of free $\beta$ -subunit of human chorionic	
	Measurement range: 0-200 ng/ml.	
	Sensitivity: 1 ng/ml.	
D-4160	<b>PAPP-A – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of pregnancy-associated plasma protein A	
	Measurement range: 0-10000 mIU/l.	
	Sensitivity: 20 mIU/l.	
	<b>Prolactin – EIA - BEST</b>	

X-3960	Enzyme immunoassay kit for the quantitative determination of prolactin.	12x8
	Measurement range: 0-3000 mIU/l.	
	Sensitivity: 15 mIU/l.	
T-8456	<b>AFP – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of	
	$\alpha$ -Fetoprotein.	
	Measurement range: 0-400 IU/ml.	
T-8468	<b>TBG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of trophoblastic $\beta$ -1-glycoprotein.	
	Measurement range: 0-200 ng/ml.	
	Sensitivity: 1.5 ng/ml.	
T-8552	<b>Ferritin – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of trophoblastic $\beta$ -1-glycoprotein.	
	ferritin.	
	Measurement range: 0-500 ng/ml.	
	<b>Sensitivity: 2.5 ng/ml.</b>	
<b>CANDIDIASIS</b>		
D-4652	<b>Candida – IgG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Candida</i> fungi.	
D-4654	<b>Candida – IgM – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgM against <i>Candida</i> fungi.	
D-4656	<b>Candida – IgA – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgA against <i>Candida</i> fungi.	
<b>ASPERGILLOSIS</b>		
D-4752	<b>Aspergill – IgG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the detection of IgG against <i>Aspergillus</i> fungi.	
<b>ANAEMIA</b>		
T-8552	<b>Ferritin – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of ferritin.	
	Measurement range: 0-500 ng/ml.	
	Sensitivity: 2,5 ng/ml.	
A-8776	<b>Erythropoietin – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of erythropoietin.	
	Measurement range: 0-200 mIU/ml.	
	Sensitivity: 0,5 mIU/ml.	
<b>TUMOR MARKERS</b>		

D-4154	<b>HCG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of human chorionic gonadotropin.	
	Measurement range: 0-500 IU/l.	
	Sensitivity: 2 IU/l.	
D-4158	<b>Free <math>\beta</math>-HCG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of free $\beta$ -subunit of human chorionic gonadotropin.	
	Measurement range: 0-200 ng/ml.	
	Sensitivity: 1 ng/ml.	
T-8454	<b>CEA – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of carcinoembryonal antigen.	
	Measurement range: 0-80 ng/ml.	
	Sensitivity: 0.3 ng/ml.	
T-8456	<b>AFP - EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of $\alpha$ -fetoprotein.	
	Measurement range: 0-400 IU/ml.	
	Sensitivity: 2.5 IU/ml.	
T-8453	<b>PSA total – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of total prostate specific antigen (PSA).	
	Measurement range: 0-5 ng/ml.	
	Sensitivity: 0.01 ng/ml.	
T-8460	<b>PSA free – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of free prostate specific antigen (PSA).	
	Measurement range: 0-5 ng/ml.	
	Sensitivity: 0.1 ng/ml.	
T-8466	<b>CA – 125 – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of CA-125.	
	Measurement range: 0-400 IU/ml.	
	Sensitivity: 1.5 IU/ml.	
T-8468	<b>TPBG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of trophoblastic $\beta$ -1-glycoprotein.	
	Measurement range: 0-200 ng/ml.	
	Sensitivity: 1.5 ng/ml.	
T-8470	<b>CA 19-9 – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative	
	Measurement range: 0-300 IU/ml.	
	Sensitivity: 2 IU/ml.	
T-8472	<b>CA 15-3 – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative	
	Measurement range: 0-250 IU/ml.	
	Sensitivity: 1,5 IU/ml.	
	<b>Ferritin-EIA-BEST</b>	

T-8552	Enzyme immunoassay kit for the quantitative determination of human ferritin.	12x8
	Measurement range: 0-500 ng/ml	
	Sensitivity: 2.5 ng/ml.	
D-3762	<b>Pepsinogen I – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of pepsinogen I.	
	Measurement range: 0-200 mkg/l. Sensitivity: 1 mkg/l.	
D-3764	<b>Pepsinogen II – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of pepsinogen II.	
	Measurement range: 0-50 mkg/l. Sensitivity: 1 mkg/l.	
<b>ACUTE PHASE PROTEINS</b>		
A-9002	<b>CRP – EIA – BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of C-reactive protein (CRP)	
	Measurement range: 0-10 IU/l Sensitivity: 0.05 IU/l	
A-9004	<b>Procalcitonin – EIA –BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of Procalcitonin.	
	Measurement range: 0-12,8 ng/ml. Sensitivity: 0,04 ng/ml.	
<b>CARDIAC MARKERS</b>		
A-9002	<b>CRP – EIA – BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of C-reactive protein (CRP).	
	Measurement range: 0-10 IU/l. Sensitivity: 0.05 IU/l.	
A-9102	<b>NTproBNP – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of N-terminal fragment of brain natriuretic	
	Measurement range: 0-2000 pg/ml. Sensitivity: 20 pg/ml.	
A-9104	<b>FABP – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of fatty-acid-binding protein.	
	Measurement range: 0-15 ng/ml. Sensitivity: 0.05 ng/ml.	
A-9106	<b>Troponin I – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of troponin I.	
	Measurement range: 0-6 ng/ml. Sensitivity: 0.02 ng/ml.	
	<b>Myoglobin – EIA - BEST</b>	

A-9108	Enzyme immunoassay kit for the quantitative determination of myoglobin.	12x8
	Measurement range: 0-1000 ng/ml.	
	Sensitivity: 4 ng/ml.	
<b>TROMBOSIS</b>		
D-9120	<b>D-dimer - EIA – BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of D-dimer in human plasma.	
	Measurement range: 0-1500 ng/ml	
	Sensitivity: 10 ng/ml	
<b>SYSTEMIC AND AUTOIMMUNE DISEASES</b>		
D-3754	<b>IgG – Gliadin – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of IgG against gliadin.	
	Measurement range: 0-100 IU/ml.	
	Sensitivity: 2.5 IU/ml.	
D-3756	<b>IgA – Gliadin – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of IgA against gliadin.	
	Measurement range: 0-100 IU/ml.	
	Sensitivity: 2.5 IU/ml.	
D-3758	<b>IgA – transglutaminase – EIA - BEST</b>	6x16
	Enzyme immunoassay kit for the quantitative determination of IgA auto-antibodies against tissue	
	Measurement range: 2.5-100 U/ml.	
	Sensitivity: 2.5 U/ml.	
X-3958	<b>Anti – TG – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of total antibodies against thyroglobulin.	
	Measurement range: 0-2000 IU/ml.	
	Sensitivity: 10 IU/ml.	
X-3968	<b>Anti – TPO – EIA – BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of total antibodies against thyroid	
	Measurement range: 0-1000 IU/ml.	
	Sensitivity: 5 IU/ml.	
A-8652	<b>VectoRheumatoid factor M-class</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of M-class Rheumatoid factor.	
A-8654	<b>VectoRheumatoid factor total</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of total Rheumatoid factor.	
A-8656	<b>Vecto-dsDNA–IgG</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of IgG autoimmune antibodies against double-stranded DNA.	
A-8658	<b>Vecto - ssDNA – IgG</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of IgG autoimmune antibodies against single-stranded DNA.	



A-8660	<b>IgE total – EIA – BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of total IgE.	
	Measurement range: 0-800 IU/ml.	
	Sensitivity: 2.5 IU/ml.	
<b>CONTROL OF THE HUMORAL IMMUNE STATUS</b>		
A-8660	<b>IgE total – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of total IgE.	
	Measurement range: 0-800 IU/ml.	
	Sensitivity: 2.5 IU/ml.	
A-8662	<b>IgG total – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative	
	Measurement range: 0-24 mg/ml.	
	Sensitivity: 0.2 mg/ml.	
A-8664	<b>IgM total – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative	
	Measurement range: 0-3.2 mg/ml.	
	Sensitivity: 0.032 mg/ml.	
A-8666	<b>IgA total – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative	
	Measurement range: 0-4.2 mg/ml.	
	Sensitivity: 0.021 mg/ml.	
A-8674	<b>Immunoscreen G, M, A – EIA - BEST</b>	32x3
	Enzyme immunoassay kit for the determination of total IgA, IgM, IgG concentration.	
A-8668	<b>IgA secretory – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of secretory IgA (sIgA).	
	Measurement range: 0-20 mg/l.	
	Sensitivity: 0.35 mg/l.	
<b>CYTOKINES</b>		
A-8752	<b><math>\gamma</math>-Interferon – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of $\gamma$ -interferon in human biological fluids	
	Measurement range: 0-1000 pg/ml.	
	Sensitivity: 5 pg/ml.	
A-8754	<b>Interleukin-4 – EIA - BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of interleukin-4 in human biological fluids	
	Measurement range: 0-100 pg/ml.	
	Sensitivity: 0.4 pg/ml.	
A-8756	<b><math>\alpha</math> -TNF - EIA-BEST</b>	12x8
	Enzyme immunoassay kit for the quantitative determination of tumor necrosis factor - $\alpha$ in human	
	Measurement range: 0-250 pg/ml.	
	Sensitivity: 2 pg/ml.	
	<b><math>\alpha</math>-Interferon – EIA - BEST</b>	

A-8758	Enzyme immunoassay kit for the quantitative determination of $\alpha$ -interferon in human biological fluids Measurement range: 0-500 pg/ml. Sensitivity: 5 pg/ml.	12x8
A-8760	<b><math>\alpha</math>-Interferon - autoimmune antibodies – EIA - BEST</b> Enzyme immunoassay kit for the quantitative determination of autoimmune antibodies against $\alpha$ - Measurement range: 0-100 ng/ml. Sensitivity: 0.4 ng/ml.	12x8
A-8762	<b>Interleukin-8 – EIA - BEST</b> Enzyme immunoassay kit for the quantitative determination of interleukin-8 in human biological fluids Measurement range: 0-250 pg/ml. Sensitivity: 2 pg/ml.	12x8
A-8766	<b>Interleukin-1 <math>\beta</math> – EIA - BEST</b> Enzyme immunoassay kit for the quantitative determination of interleukin-1 beta in human biological Measurement range: 0-250 pg/ml. Sensitivity: 1 pg/ml.	12x8
A-8768	<b>Interleukin-6 – EIA - BEST</b> Enzyme immunoassay kit for the quantitative determination of interleukin-6 in human biological fluids and culture medium Measurement range: 0-300 pg/ml. Sensitivity: 0.5 pg/ml.	12x8
A-8770	<b>Interleukin-18 – EIA - BEST</b> Enzyme immunoassay kit for the quantitative determination of interleukin-18 in human biological Measurement range: 0-1000 pg/ml. Sensitivity: 5 pg/ml.	12x8
A-8772	<b>Interleukin-2 – EIA - BEST</b> Enzyme immunoassay kit for the quantitative determination of interleukin-2 in human biological fluids Measurement range: 0-500 pg/ml. Sensitivity: 2 pg/ml.	12x8
A-8774	<b>Interleukin-10 – EIA - BEST</b> Enzyme immunoassay kit for the quantitative determination of interleukin-10 in human biological Measurement range: 0-500 pg/ml. Sensitivity: 1 pg/ml.	12x8
A-8776	<b>Erythropoietin – EIA - BEST</b> Enzyme immunoassay kit for the quantitative determination of erythropoietin in human biological Measurement range: 0-200 mIU/ml. Sensitivity: 0.5 mIU/ml.	12x8